

Confidential  
TCLVI-EO7

C.F. Tsai  
Formosa Plastics Corporation  
1 Hsin-Hwa 1st Road  
Lin-Yuan Village  
Kaohsiung Hsien,  
Taiwan

Lab No. 07T\_41102\_02  
P.O. No. 75708007  
Test Facility: NAMSA  
6750 Wales Road  
Northwood, OH 43619

CERTIFICATE OF COMPLIANCE  
USP BIOLOGICAL REACTIVITY TESTS, *IN VIVO*

USP PLASTIC CLASS VI

Test Article: Polypropylene Pellets, YOUNGSOX 5350T  
ID No. See Test Article

USP Systemic Toxicity Study in the Mouse: The test article was prepared as indicated below and injected into mice. The saline, alcohol in saline, polyethylene glycol 400 and sesame oil extracts did not produce a significantly greater systemic reaction than the blank extractants.

USP Intracutaneous Toxicity Study in the Rabbit: The test article was prepared as indicated below and injected intracutaneously into rabbits. The saline, alcohol in saline, polyethylene glycol 400 and sesame oil extracts did not produce a significantly greater tissue reaction than the blank extractants.


USP Muscle Implantation Study in the Rabbit: The macroscopic reaction of the test article, implanted in rabbit muscle for 1 week, was not significant when compared to the USP negative control plastic.

The test article was prepared at a ratio of 4 g:20 ml and extracted at 70°C for 24 hours. The test article extracts met the requirements of a USP Plastic Class VI.

mdo Date Completed

8-17-07

Approved By



Christina L. Oval, B.A.  
Technical Writer